

LIMITED WARRANTY

Den-Mat will replace or refund the purchase price of any of its products that are proven to be defective within 30 days of the purchase date. Replacement of defective goods, or refund of the purchase price shall be the exclusive remedy for the user. Den-Mat will not be liable for any economic, incidental, or consequential loss or damage that arises out of the use of, or the inability to use its products.

This limited warranty is in lieu of all other warranties, expressed or implied, and shall be void if the product is improperly stored or used. There are no implied warranties of merchantability, fitness for a particular purpose, or otherwise. Before using this product, the user shall determine whether it is suitable for the intended use and the user shall assume all risk and liability associated therewith.

DENMAT CUSTOMER SUPPORT

U.S. and Canada 800-433-6628
U.S. and Canada Fax Orders and Inquiries..... 800-233-6628
International Inquiries 805-347-7990
International Fax Orders and Inquiries: 805-347-7983
Website denmat.com



Tenure S Value Kit
Bond Enhancer




REF 031145100

CONTENTS:

- Tenure S Bond Enhancer, 18 ml **CE 0044**
- Instruction card **(NM-D)**



 Den-Mat Holdings, LLC
1017 W. Central Avenue
Lompoc, CA 93436 USA

EC	REP	MDSS GmbH Schiffgraben 41 30175 Hannover, Germany
----	-----	---

Symbols Glossary available at: denmat.com/symbols

Tenure® S is an easy-to-use, one-step, light-cure, low-viscosity, bond enhancer that can also be used as a wetting resin in composite restorations

CAUTION: Wear protective gloves while using this product.

CAUTION: Wear eye protection while using this product.

CAUTION: Tenure® S has not been studied in children, pregnant or breast-feeding women.

DIRECTIONS

1. Apply Tenure A and B per instructions.
2. Brush one coat of Tenure S (prior to the composite placement).
3. Light-cure, starting with:
 - 5 seconds — Sapphire® Plus Plasma Arc Curing Light
 - 10 seconds — Flashlite® LED Curing Lights
 - 30 seconds — Conventional curing light (e.g., halogen)

Contraindications

Patients allergic or hypersensitive to methacrylate monomers.

STORAGE

Do not expose to temperatures exceeding 77° F (25° C).

Do not expose to direct sunlight.

Do not refrigerate.

Do not freeze.

RELATED DENMAT PRODUCTS

Description

Product Number

Tenure Quik® with Fluoride	031145700
Tenure® Multi-Purpose Bonding System.....	031146000
Tenure Uni-Bond® with Gloss-N-Seal®.....	030411900

SAFETY DATA SHEETS AVAILABLE AT denmat.com